


REVISIONS					
REV	ECO	DESCRIPTIONS	DWN	CKD	APVD
A	✓	RELEASE TO ECO CONTROL	LML 11/18/99	MPT 11/29/99	PMF 11/30/99
B	03439	UPDATED PER ECO	RWN 01/27/00		

HAZARDOUS LOCATION  
CLASS I, DIVISION 1, GROUP A,B,C,D

NON-HAZARDOUS OR HAZARDOUS  
CLASS I, DIVISION 2, GROUP B,C,D LOCATION  
ASSOCIATED EQUIPMENT/ APPAREILLAGE CONNEXE

NON-HAZARDOUS LOCATION



CSA CONTROLLED DOCUMENT DO NOT ALTER WITHOUT CSA AUTHORIZATION

- NOTES:
- A SINGLE DELTA F OXYGEN SENSOR AND SINGLE TF/M SERIES PROBE MAY CONNECT TO THE DESIGNATED TERMINALS OF THE MMS 3.
  - "WARNING - EXPLOSION HAZARD - SUBSTITUTION OF COMPONENTS MAY IMPAIR SUITABILITY FOR CLASS I, DIVISION 2" and "AVERTISSEMENT - RISQUE D'EXPLOSION - LA SUBSTITUTION DE COMPOSANTS PEUT RENDRE CE MATERIAL INACCEPTABLE POUR LES EMPACEMENTS DE CLASSE I, DIVISION 2."
- "WARNING - EXPLOSION HAZARD - DO NOT REPLACE FUSE UNLESS POWER HAS BEEN SWITCHED OFF OR THE AREA IS KNOWN TO BE NON-HAZARDOUS" and "AVERTISSEMENT - RISQUE D'EXPLOSION - COUPER LE COURANT OU S'ASSURER QUE L'EMPLACEMENT EST DESIGNÉ NON DANGEREUX AVANT DE REPLACER LE FUSE."
- "WARNING - EXPLOSION HAZARD - DO NOT DISCONNECT EQUIPMENT UNLESS POWER HAS BEEN SWITCHED OFF OR THE AREA IS KNOWN TO BE NON-HAZARDOUS" and "AVERTISSEMENT - RISQUE D'EXPLOSION - AVANT DE DECONNECTER L'EQUIPEMENT, COUPER LE COURANT OU S'ASSURER QUE L'EMPLACEMENT EST DESIGNÉ NON DANGEREUX."

		UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE		<b>PANAMETRICS</b> WALTHAM, MASS. 02154-3497	
		FRACTIONS ± ✓	DECIMALS .XX ± .01	TITLE	
			.XXX ± .005	IS SYSTEM DIAGRAM	
			SURFACE FINISH 125 ✓	DRAWING NUMBER	
		DRAWN LML 11/18/99	ENGINEER	SIZE <b>B</b>	752-138
		CHECKED		REV <b>B</b>	
B	03439	RN 4880		SHEET 3 OF 3	
A	✓				
REV	ECO	MODEL NO. MMS 3	SCALE FULL	DO NOT SCALE DWG	